

CAST PARTIAL DENTURE

Certificate of Authenticity

The dental restorations used for your case were constructed of FDA approved 510k materials and hand crafted in our North American Digital Processing Center. This is your assurance of quality and safety in materials and construction.

Item	Product Name	Manufacturer	Country of Origin	FDA 510k	Description
Metal Framework	Wironit®LA	Bego	Germany	A150801 (prior to 1976 Preamendment)	Co 63.5%, Cr 29%, Mo 5%, <2% Si, Mn, N, C, Ta
Acrylic	Lucitone 199®	Dentsply	Germany	K011560	Polymethylmethacrylate and Methyl Methacrylate; <10% of ethylene glycol dimethacrylate
Standard Acrylic Teeth	HK Artic	Haraeus Kulser	Germany	K033628	Poly(methacrylate)
Premium Denture Teeth	SR-Vivodent & SR-Orthotype	Ivoclar	Germany	K864707	Synthetic Polymer teeth based on polymethylmethacrylate



Heikas Martirosian

Heikas Martirosian, CDT
Director of Technical Training